


INTERPRETATION GUIDE

Bowie-Dick Test Sheet - BD 110

UNEXPOSED

Site Luogo Localisation	Operator Operatore Operador	Date Data Fecha	Department Reparto Departamento
Steri-Pak[™] LF REF BD110 LEAD FREE BOWIE-DICK TEST SHEET			
			
LOT 040701 Steam/Vapore/Vapor 2007-07 EN 867-3 CLASS B 134°C - 3½ MIN			
Supervisor Supervisore Supervisor	Machine No. Autoclave	Cycle No. Ciclo	Result Risultato Resultado

CORRECT COLOR CHANGE

Site Luogo Localisation	Operator Operatore Operador	Date Data Fecha	Department Reparto Departamento
Steri-Pak[™] LF REF BD110 LEAD FREE BOWIE-DICK TEST SHEET			
			
LOT 040701 Steam/Vapore/Vapor 2007-07 EN 867-3 CLASS B 134°C - 3½ MIN			
Supervisor Supervisore Supervisor	Machine No. Autoclave	Cycle No. Ciclo	Result Risultato Resultado

INSUFFICIENT

Site Luogo Localisation	Operator Operatore Operador	Date Data Fecha	Department Reparto Departamento
Steri-Pak[™] LF REF BD110 LEAD FREE BOWIE-DICK TEST SHEET			
			
LOT 040701 Steam/Vapore/Vapor 2007-07 EN 867-3 CLASS B 134°C - 3½ MIN			
Supervisor Supervisore Supervisor	Machine No. Autoclave	Cycle No. Ciclo	Result Risultato Resultado



74 Inverness Drive East - Englewood, Colorado 80112

Phone: 303-660-4201 - Fax: 303-660-4213 - E-mail: info@steritecproducts.com -
Website: www.steritecproducts.com

MK-040 Rev. B